

*Ron Wilcox
Will you provide the
info ~~of~~ required here,
JA*

MSG REF: 111111 DTG: 04/19/83 10:35
TO: JOHNNY ADLER PCB
CC: MIKE SINGH BALL
FM: JEAN VOISIN EBRA

RE: RS232 CARD

JOHNNY:

WE KNOW THAT UP TO TWO RS232 CARDS COULD BE INSTALLED INSIDE A SINGLE PERIPHERAL EXPANSION BOX.

UNFORTUNATELY WE DONT KNOW, IN EUROPE, HOW TO MODIFY ONE OF THE TWO CARDS TO BE THE SECOND ONE.

COULD YOU TELL ME WHAT NEED TO BE DONE AND HOW YOU PROCEED IN AMERICA.

THANKS IN ADVANCE AND REGARDS

JEAN VOISIN

ECG QUALITY AND SERVICE MGR

TO: Jean Voisin

FR: Ron Wilcox

RE: 2ND RS232

Attached layout and schematic should allow proper mod.

Regards,
Ron Wilcox

TIF

TO INSTALL THE SECOND RS232 IN THE EXPANSION BOX YOU MUST:

1. REMOVE R5 = 100Ω
2. REINSTALL R5 IN THE HOLE PATTERN DIRECTLY BELOW THE ORIGINAL INSTALLATION

HP 1039309 REV G
MARKING

THE END RESULT IS TO PULL UP U7-19 TO 5V SEE SCHEMATIC



